

Lab name: SRI Instruments
 Client: SRI R&D
 Client ID: N10857
 Analysis date: 10/20/2019 13:58:23
 Method: 1ul Heated Inj=200C
 Description: FID Detector
 Column: 30MXT502.2 1 mic
 Carrier: H2 from BIHG
 Integration: Peak sens=95.0 Base sens=60.0 Min area= 100.00 Standard= 4.000 Sample=100.000 Tangents=off
 Data file: Cannabis165.chr ()
 Sample: terp and cannabinoid mix
 Comments: Injector=200 FID=300

Temperature program:

| Init temp | Hold | Ramp | Final temp |
|-----------|-------|--------|------------|
| 50.00 | 0.000 | 10.000 | 70.00 |
| 70.00 | 4.000 | 30.000 | 300.00 |
| 300.00 | 3.000 | 0.000 | 300.00 |
| 40.00 | 3.000 | 10.000 | 50.00 |

